

CONDUCTIVITY STANDARD

CONDUCTIVITY STANDARD

85 $\mu\text{S}/\text{CM}$

Conductivity Value: 85 $\mu\text{S}/\text{cm} \pm 1\%$ @ 25°C

Volume: 500 mL

Product Code: CDS-PW01-500

Appearance: Red

Application: Prepared from high-purity potassium chloride (KCl), ideal for calibrating low-range conductivity meters in laboratory, industrial, and environmental testing.



CONDUCTIVITY STANDARD

1000 $\mu\text{S}/\text{CM}$

Conductivity Value: 1000 $\mu\text{S}/\text{cm} \pm 1\%$ @ 25°C

Volume: 500 mL

Product Code: CDS-PW02-500

Appearance: Red

Application: Prepared from high-purity potassium chloride (KCl), ideal for calibrating low-range conductivity meters in laboratory, industrial, and environmental testing.



CONDUCTIVITY STANDARD

1413 $\mu\text{S}/\text{CM}$

Conductivity Value: 1413 $\mu\text{S}/\text{cm} \pm 1\%$ @ 25°C

Volume: 500 mL

Product Code: CDS-PW03-500

Appearance: Red

Application: Widely used reference standard prepared with KCl, perfect for precise calibration of conductivity meters in industrial, chemical, and environmental applications..



ERBIL OFFICE

597 Italian Village 2, Erbil 44001, IRAQ
+964 (782) 373 7944
+964 (750) 857 3942
info@slc-erbil.com

BAGHDAD OFFICE

Business Avenue Complex - Office D11, Abu
afer Al-Mansour Square, Baghdad 10081, IRAQ
+964 (781) 891 0575
info@smartlabcompany.com



Ur SYSTEM

Part Of Smart Lab Co.